



# C-reactive Peptide

Version: 1  
 Edited by: Jason Kim

*(note that the following list should be linked to the appropriate location.)*

- [Summary](#)
- [Reagents and Materials](#)
- [Protocol](#)
- [Reagent Preparation](#)
- [Reagent 1](#)
- [Reagent 2](#)
- [Reagent 3](#)

**Summary:** *(This area will include a brief description of what the protocol is used for and why someone would need to use it.)*

This experiment involves a spectrophotometric measurement using Roche Cobas Clinical Chemistry Analyzer. Serum levels of C-reactive peptide reflect systemic inflammation and may be altered in obesity.

**Reagents and Materials:** *(This should be a comprehensive list of stock solutions and material. The reagent list for the stock solutions is included in the reagent preparation area that is included at the end of this SOP.)*

Reagent/Material	Vendor	Stock Number
C-reactive protein (latex)	Roche	05401615 190
Calibrator f.a.s. Proteins		11355279 160
Precinorm Protein		10557897 160
Precipath Protein		11333127 160
NaCl Diluent 9 %		04774230 190
Chimneys		11930630 001
Cleaner		04774248 190
Micro Sample cups		11406680 001
NERL High Quality Water		Fisher

## Protocol:

**Notes:**

- ✓ Try to use freshly prepared serum and plasma samples for this assay.
- ✓ No dilution or treatment of the sample is required, but plasma samples should be centrifuged to remove any fibrin/fibrinogen clumps.

- ✓ Samples should be stored at 2-8°C for 24 hours prior to analysis. For longer periods, store samples at -70°C, and avoid repeated freeze/thaw cycles.
  - ✓ A 50 µl dead volume is required in addition to sample volume for multi-protein analysis (typically 1-5 µl).
1. Perform daily quality control assessment of instrumentation before analysis.
  2. Load each sample into a specialized micro-sample cup for the clinical chemistry analyzer.
  3. Select C-reactive Peptide test on display and run the analysis.
  4. Collect and analyze the data.